

NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance for Weighing and Measuring Devices

For: Indicating Element **Digital Electronic** Model: Counterpart n<sub>max</sub>: 5 000 Accuracy Class: III / III L Submitted By:

Rice Lake Weighing Systems, Inc. 230 West Coleman St Rice Lake, WI 54868 Tel: 715-234-9171 Ext. 5322 Fax: 715-234-6967 Contact: Paul A. Lewis, Sr. Email: plewis@ricelake.com Web site: www.ricelake.com

## **Standard Features and Options**

## **Standard Features:**

- Semi- automatic (push-button) Zero (SAZSM)
- Semi-Automatic (push-button) Tare
- Programmable Tare
- LCD Display
- Center of Zero Annunciator
- Alphanumeric Keypad
- Printer Interface
- Unit Conversion (kg, g, lb, oz)
- Linear Calibration Points (5)
- AC Power
- Two RS-232 Ports
- 20 mA Current Loop
- Category 2 Audit Trail

## **Options:**

- Dual Channel
- DC Power (Battery) Supply
- Ethernet
- USB
- AC/DC Adapter
- Automatic Zero Tracking (AZT)
- Keyboard Tare
- Gross/Net Display
- Sleep Mode
- Alphanumeric Display
- Remote Calibration Wireless Communications

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Stephen Benjamin Chairman, NCWM, Inc.

Kurt Floren

Committee Chair, National Type Evaluation Program Committee Issued: November 19, 2012

#### 1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



# **Rice Lake Weighing Systems, Inc.**

Indicating Element / Counterpart

Application: General purpose scales. The indicator is approved for use with compatible NTEP certified weighing and printing elements.

**Identification:** The identification badge is located on the back of the indicator.

Sealing: The indicator can be sealed with a lead and wire seal threaded through two drill head screws in the left side of the indicator.

This unit includes a Category 2 Audit Trail. The device may still be sealed with a physical seal as described above or the inspector may record the audit trail event counters at the time of test. To view the CALIB Counter press the [MENU SETUP] key, then press the Down Arrow [GROSS NET] key, then press the Right Arrow [PRINT] key. This will display the CALIB Counter. To view the CONFIG Counter press the Up Arrow [ZERO] key, then press the Right Arrow [PRINT] key, then press the Down Arrow [GROSS NET] key. This will display the CONFIG Counter. To exit, press the [SAVE EXIT] soft key at the bottom left corner of the display.

<u>Test Conditions</u>: This emphasis of this evaluation was on the device design, operation, performance, and compliance with influence factor requirements. One Counterpart indicator was submitted for evaluation and was interfaced with a load cell simulator and a scale base at the same time in order to demonstrate dual scale mode functionality. The indicator was tested over a temperature range of -10  $^{\circ}$ C to 40  $^{\circ}$ C (14  $^{\circ}$ F to 104  $^{\circ}$ F). Additionally, tests were conducted using power supplies of 100 and 240 VAC, then 9 and 15 VDC. The wireless printing features were tested in conformance to NCWM Publication 14 requirements.

## Evaluated By: J. Morrison (OH)

**Type Evaluation Criteria Used:** NIST Handbook 44 Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices, 2012 Edition. NCWM Publication 14 Weighing Devices, 2012 Edition.

<u>Conclusion</u>: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: J. Truex (NCWM)

## Example of Device:



Counterpart