



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

Non-Computing Scale
Bench Scale, Electronic
Model: AP Series
 n_{\max} : 4 000
 e_{\min} : (see below)
Capacity: (see below)
Platform: (see below)
Accuracy Class: III

***Submitted By: Contact Info. Updated November 2019**

Cardinal Scale Manufacturing Company
102 East Daugherty Street
Webb City, MO 64870
Tel: 417-673-4631 x 211
Fax: 417-673-5001
Contact: Eric Golden
Email: egolden@cardet.com
Web site: www.cardet.com


Standard Features and Options**Standard Features:**


- Semi-automatic, Automatic and Initial Zero Setting Mechanisms
- Semi-automatic Tare
- Battery Power Supply with Sleep and Automatic Shut-off Modes
- Gross/Net Display
- Pound/kg Conversion

Model	Capacity and Division (d)	Platform Size (in)
AP-4K	4000 g x 1 g	6.75 x 5.25
AP-4KD	4000 g x 1 g	6.75 x 5.25
AP-4KDEX	4000 g x 1 g	6.75 x 5.25
AP-4KEX	4000 g x 1 g	6.75 x 5.25
AP-6	99.95 oz x 0.05 oz	6.75 x 5.25
AP-6EX	99.95 oz x 0.05 oz	6.75 x 5.25
AP-6PH	99.9 oz x 0.1 oz	6.75 x 5.25
AP-8	8.000 lb x 0.002 lb	6 x 9
AP-10	9.995 lb x 0.005 lb	6.75 x 5.25
AP-10CG	9.99 lb x 0.01 lb	6.75 x 5.25
AP-10K	9995 g x 5 g	6.75 x 5.25
AP-10KEX	9995 g x 5 g	6.75 x 5.25

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages. *Editorial changes, not affecting the type or metrological content, corrected this certificate.


Tim Tyson
Chairman, NCWM, Inc.


Randy Jennings
Chairman, National Type Evaluation Program Committee
Issued: October 27, 2010

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Cardinal Scale Manufacturing Co.

Non-Computing Scale / AP Series

Application: General purpose weighing.

Identification: The identification badge is on the side of the scale.

Sealing: The device can be sealed with a physical seal.

Test Conditions: This certificate supersedes Certificate of Conformance (CC) No. 95-087 and is issued without additional testing to correct the omission of the Model AP-8 and the declaration of platform sizes. The test conditions for the previous certificate are listed below for reference.

Certificate of Conformance Number 95-087: The emphasis of this evaluation was on device design, performance, and compliance with influence factors. The Models AP-10K, AP-4K, and AP-8 were tested. These devices were tested over a temperature range of -10 °C to 40 °C. In addition, a permanence test was conducted on the AP-8 during which the device was tested periodically. The device was also tested over a voltage range of 100 VAC to 130 VAC.

Evaluated By: Milton G. Smith (Canada, LMB)

Type Evaluation Criteria Used: NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 1996. NCWM, Publication 14: Weighing Devices, 1996.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: C. V. Cotsoradis (NIST) (95-087); L. T. Sebring (NIST) (95-087A1)