



Safety Warning

Technical Document 0069

September 12th 2013

Model SEK Checkweighing Scale Safety Warning

This warning refers to the light tower accessory and not to the main weighing machine.

Please Take Action Immediately

Please ensure that all existing SEK light towers are removed from service immediately. Intelligent Weighing Technology has issued a replacement light towers to replace the unsafe unit. Please replace the existing SEK light tower immediately and dispose of safely.

Please note that the SEK 120 VAC wiring has a Green Ground wire with connecting loop. This wire is meant to be connected to the L shaped metal bracket as the attached photograph shows.

Please seek the services of a qualified electrician to perform this task if you are unsure of your capabilities.

Note!

The model SEK is not designed for use in areas where water, or other contamination, may result in compromising the electrical system safety. Please ensure that all SEK units are not used in these environments.